

Member of the FM Global Group

FM Approvals 1151 Boston Providence Turnpike P.O. Box 9102 Norwood, MA 02062 USA T: 781 762 4300 F: 781-762-9375 www.fmapprovals.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

RL20000ST Load Cell.

IS / I,II,III /1 /ABCDEFG/T4 Ta=40°C — 66147; Entity*; NI /I /2 /ABCD/T4 Ta=40°C; S / II,III /2 /FG/T4 Ta=40°C

*Refer to Dwg. 66147 for entity parameters

Equipment Ratings:

Intrinsically safe for Class I, II, III, Division 1, Groups A,B,C,D,E,F and G,T4 Ta=40°C hazardous (classified) locations. When installed in accordance with Rice Lake Weighing Systems Control Drawing Number 66147.

Nonincendive for Class I, Division 2, Groups A,B,C,D, Suitable for Class II, III, Groups F and G,T4 Ta=40°C hazardous (classified) locations.

FM Approved for:

Rice Lake Weighing Systems 230 W. Coleman Rice Lake, WI 54868



This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

Class 3600	1998
Class 3610	2010
Class 3611	2004
Class 3810	1989

Original Project ID: 3015962

Approval Granted: March 31, 2003

Subsequent Revision Reports / Date Approval Amended

Report Number

Date

Report Number

Date

3042054

February 1, 2012

FM Approvals LLC

J. E. Marquedant

Group Manager, Electrical

E. Marquestint

1 February 2012

Date