



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Indicating Element
Digital Electronic
Model: 480 & 482 Series
 n_{max} : 10 000
Accuracy Class: III / III L

Submitted By:

Rice Lake Weighing Systems, Inc.
230 West Coleman St
Rice Lake, WI 54868
Tel: 715-234-9171 Ext. 5322
Fax: 715-234-6967
Contact: Paul A. Lewis, Sr.
Email: plewis@ricelake.com
Web site: www.ricelake.com

Standard Features and Options

Model	Screen Type	Enhanced Features	Suffix XY (X = Enclosure Type and Y = Power Input)
480-XY	LED	No, 7 keys	X = 2 (Stainless Steel) Y = A (AC Voltage)
480Plus-XY	LED	Yes, 19 keys	X = 2 (Stainless Steel) Y = A (AC Voltage)
482-XY	LCD	No, 7 keys	X = 2 (Stainless Steel) Y = A (AC Voltage)
482Plus-XY	LCD	Yes, 19 keys	X = 2 (Stainless Steel) Y = A (AC Voltage)

- Semi-Automatic (push-button) Zero (SAZSM)
- Automatic Zero Tracking (AZT)
- Initial Zero Setting Mechanism (IZSM)
- Semi-Automatic (push-button) Tare
- Keyboard Tare / Programmable Tare / Tare Save Key
- Gross/Net Display / LED Display
- LED Display (Model 480) / LCD Display (Model 482)
- Annunciators (Units, Gross, Net, Motion, Center of Zero, Tare, Present Tare)
- Printer Interface
- Unit Conversion (kg, g, lb,oz, Ton Metric, Short Ton)
- AC Power / DC Power
- Local/Remote Operation
- Category 2 Audit Trail / Remote Calibration / 2 Point Calibration
- 20 mA Current Loop / RS-232 Ports (2) / USB / Ethernet TCP/IP
- Wireless Communications
- Sleep Mode

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

Ronald Hayes
Chairman, NCWM, Inc.

John Gaccione
Committee Chair, National Type Evaluation Program Committee
Issued: April 9, 2015

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



Rice Lake Weighing Systems, Inc.
Indicating Element / 480 & 482 Series

Application: A general purpose indicating element to be interfaced with an approved and compatible weighing element.

Identification: The identification badge is located on the side of the indicator.

Sealing: The indicator can be sealed with a lead and wire seal threaded through three drill head screws in the left side of the indicator.

This unit includes a Category 2 Audit Trail. The device may still be sealed with a physical seal as described above or the inspector may record the audit trail event counters at the time of test. To view the Calibration and Configuration Counters press the [MENU] key, then press the Left Arrow [UNIT] key, then press the Down Arrow [GROSS/NET] key twice. This will display "CALib". To view the Calibration Counter press the Down Arrow [GROSS/NET] key twice. To view the Configuration Counter press the Up Arrow [ZERO] key twice, then press the Right Arrow [PRINT] key. This will display "ConFiG". To view the Configuration Counter press the Down Arrow [GROSS/NET] key twice. To exit, press the Up Arrow [ZERO] key four times then press the Right Arrow [PRINT] key, then press the UP Arrow [ZERO] key.

Test Conditions: This Certificate supersedes Certificate of Conformance Number 12-123A1 and is issued to correct omissions from the previous certificate and add the DC power and sleep mode features. Both the DC power and sleep mode features were evaluated by NTEP in the NTEP laboratory. Previous test conditions are listed below for reference

Certificate of Conformance Number 12-123A1: This Certificate supersedes Certificate of Conformance Number 12-123 and is issued to add models 480Plus-XY, 482-XY, and 482Plus-XY. For the purpose of this test, one 480Plus-2A, one 482-2A, and one 482Plus-2A were submitted for evaluation. The emphasis of this evaluation was on the device design, operation, performance, and compliance with influence factor requirements. The 482Plus-2A was interfaced with a load cell simulator and it was tested over a temperature range of -10 °C to 40 °C (14 °F to 104 °F). Additionally, tests were conducted using power supplies of 115 and 230 VAC.

Certificate of Conformance Number 12-123: The emphasis of this evaluation was on the device design, operation, performance, and compliance with influence factor requirements. Two 480-2A indicators were submitted for evaluation. One was interfaced with a load cell simulator. The indicator was tested over a temperature range of -10 °C to 40 °C (14 °F to 104 °F). Additionally, tests were conducted using power supplies of 115 and 230 VAC. A scale base was connected to the 480-2A to conduct the discrimination test. Local / Remote operation was tested with two 480-2A's using wireless communication. Two 720i were setup as a Local / Remote operation using wireless communication to test the wireless communication per Pub 14.

Evaluated By: C. Harris (OH) 12-123; J. Morrison (OH) 12-123A1, 12-123A2

Type Evaluation Criteria Used: *NIST Handbook 44 Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices*, 2014 Edition. *NCWM Publication 14 Measuring Devices*, 2014 Edition.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: J. Truex (NCWM) 12-123, 12-123A1, 12-123A2



Rice Lake Weighing Systems, Inc.
Indicating Element / 480 & 482 Series

Examples of Device:

