



Member State of OIML
United Kingdom of Great Britain
and Northern Ireland

OIML Certificate No
R60/1991-GB-95.08

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name:

National Weights and Measures Laboratory

Address:

Stanton Avenue

Teddington

Middlesex

TW11 0JZ

Person responsible:

M D Koch - Head of Type Examination

Applicant

Name:

Rice Lake Weighing Systems.

Address:

230 W. Coleman Street

Rice Lake

WI 54868

USA

Manufacturer of the certified pattern :

Sensortronics, Covina California

Identification of the certified pattern:

Load Cell Model No Rice Lake RLETB

Further characteristics see page 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60 edition 1991 (E) for accuracy class C

OIML Certificate No
R60/1991-GB-95.08

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report application no 30-95, that includes 26 pages.

The issuing authority

The CIML member

NWML

S J Bennett

S J Bennett

Date: 2nd June 1995

Date: 2nd June 1995

Characteristics:

Accuracy class	max output	input impedance	output impedance	cable type	cable length
C	3 mV/V	350 Ω	350 Ω	4 wire shielded. Shield not connected to load cell.	20 feet

Capacity E _{Max}	v _{min}	E _{min}	n _{max}
1000 lb	0.10 lb	20 lb	3000
1500 lb	0.15 lb	30 lb	3000
2000 lb	0.20 lb	40 lb	3000
2500 lb	0.25 lb	50 lb	3000
3000 lb	0.30 lb	60 lb	3000
4000 lb	0.40 lb	80 lb	3000
5000 lb	0.50 lb	100 lb	3000
10000 lb	1.00 lb	200 lb	3000
500 kg	0.05 kg	10 kg	3000
750 kg	0.08 kg	15 kg	3000
1000 kg	0.10 kg	20 kg	3000
1500 kg	0.15 kg	30 kg	3000
2000 kg	0.20 kg	40 kg	3000
3000 kg	0.30 kg	60 kg	3000
5000 kg	0.50 kg	100 kg	3000

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.