



**United Kingdom of Great Britain
and Northern Ireland**



**OIML Certificate No
R76/2006-GB1-10.02**

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **National Weights and Measures Laboratory
(Part of the National Measurement Office)**
Address: **Stanton Avenue
Teddington
Middlesex
TW11 0JZ
United Kingdom**

Person responsible: **Paul Dixon – Product Certification Manager**

Applicant

Name: **Rice Lake Weighing Systems**
Address: **230 W. Coleman Street
Rice Lake
WI 54868
USA**

Identification of the certified pattern:

Weighing indicators, as part of a non-automatic weighing instrument, designated the CW90-A; CW90-E; CW90X-A and CW90X-E.

Further characteristics see page 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML:	R76
Edition:	2006 (E)
Accuracy class:	III, IIII

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

**OIML Certificate No
R76/2006-GB1-10.02**

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated:

NWML Test reports:	TR 588	having 42 pages
	SN 1156	having 9 pages
	SN 1157	having 14 pages
Pattern Evaluation report:	P00270	having 14 pages

The issuing authority



Mr G E Stones

The CIML member



Mr P Mason

Date: 28 October 2010

Ref: T1128/0236

Characteristics: The indicating devices are designated CW90-A; CW90-E; CW90X-A and CW90X-E. They are designed to be used as part of a Class III or IIII non-automatic weighing instruments. The indicators are self-indicating, mains or DC-powered and operate as single-interval instruments.

Main features:

- Controller housed in stainless steel enclosure
- LED display with enunciators (stability, zero, gross/net, units, tare, preset tare, percentage)
- "UNDER", "ACCEPT" and "OVER" bands
- Function and navigation keys
- Numeric keypad (CW90-A and CW90-E only)
- Mains-powered (CW90-A and CW90X-A) or DC-powered (CW90-E and CW90X-E)

Devices:

- Semi-automatic zero setting ($\leq 4\%$ of Max)
- Zero-tracking ($\leq 0.5d/s$ within 4% Max)
- Subtractive semi-automatic tare balancing
- Pre-set tare
- Checkweighing
- Printing
- Direct sales to the public (if Preset Tare is disabled)

Technical data:

Power supply	100-240 VAC - 50/60 Hz 10-60 VDC
Maximum number of scale intervals	10,000
Maximum tare	- Max
Load cell excitation voltage	5 Vdc
Minimum load cell impedance	43.75 Ω
Maximum load cell impedance	2000 Ω
Minimum input voltage per verification scale interval	0.85 μ V
Measuring range minimum voltage	0 mV
Measuring range maximum voltage	70 mV
Fraction of maximum permissible error	$P_{ind} = 0.5$
Operating temperature range	- 10 °C to + 40 °C
Load cell cable (from indicator to load cell junction box) - Maximum length	276 m/mm ² (6-wire configuration) 0.5 m (4-wire configuration)

Interfaces / optional boards:

The instrument may be fitted with the following protected interfaces:

- Load cell 4 or 6-wire connection
- RS232
- 20mA Current Loop
- Ethernet TCP/P (wired, shielded cable required)
- Wireless Lan Card Wi-Fi 802.11 A/B
- USB
- Fiber optic
- Digital I/O (braided cable with added ferrite required)

Certificate History

ISSUE NO.	DATE	DESCRIPTION
R76/2006-GB1-10.02	28 October 2010	Certificate first issued.
-	-	No revisions have been issued.

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.