

Certificate of Weight Calibration (Non-accredited)

A customer requesting a non-accredited Certificate of Weight Calibration, needs proof of traceability to NIST and actual values and uncertainties. Comparisons must be made between the item being calibrated and the standard being used. The mass reference standard used for the tolerance test is essential to the traceable document.

Prior to comparison between the known standard and the item submitted for calibrating, the known standard must be sufficiently calibrated over time to produce predictable measurements.

This certificate should contain all of the data related to the calibration. After calibrating, a non-accredited Certificate of Weight Calibration is issued. Although the Certificate of Weight Calibration (non-accredited) provides traceability to NIST, it is not a NVLAP accredited document. If an accredited document is required, please refer to the Certificate of Weight Calibration (accredited) on the previous page.

RICE LAKE

Certificate of Weight Calibration

Traceable Certificate Number: 1234567 1

Contractor: 2
RICE LAKE WEIGHING SYSTEMS
230 W. COLEMAN STREET
RICE LAKE, WI 54868

Purchase Order Number: 3 PURCHASE ORDER
Client: 4
RICE LAKE WEIGHING SYSTEMS
230 W. COLEMAN STREET
RICE LAKE, WI 54868

4 Date Received: 25 Sep 2022
5 Date Calibrated: 26 Sep 2022 to 27 Sep 2022
6 Recalibration Date: 26 Sep 2023
7 NIST Certificate Number: 684/292805-19
8 If there are two NIST numbers, one may apply
9 Calibrated By: 20, 28
10 Procedure: WI05-0095 Rev. D
11 Condition of Weights: Acceptable for Calibration
12 Description of Weights: 2 mg to 100 g Polished Weights, ASTM Class 1, S/N 1234, ID# ABCD
13 Comments:

Key Notes

Finish	* Indicates the weight does not meet the finish requirements
Material	+ Indicates the weight does not meet the material requirements
New Wt	◊ Indicates new weight
Missing Wt	▲ Indicates replaced missing weight with new weight
Damaged Wt	☒ Indicates replaced damaged weight
Replaced OOT	★ Indicates replaced out of tolerance weight
OOT	☒ Indicates correction plus or minus Uncertainty greater than or equal to MPE
Magnetic Wt	★★ Indicates replaced magnetic weight
Design	☒ Indicates the weight does not meet the design or shape requirements
Repainted	● Indicates the weight was repainted after As Found obtained
Other	† See comments above

Cleaning Levels


A	Dusted with brush or cloth
B	Spot cleaned with ethyl alcohol
C	Full surface cleaned with ethyl alcohol
D	Spot cleaned with non-alcohol solvent followed by ethyl alcohol
E	Full surface cleaned with non-alcohol solvent followed by ethyl alcohol
F	No cleaning performed

Material Abbreviations

AL	Aluminum	TA	Tantalum
SS	Stainless Steel	BR	Brass
CI	Cast Iron	PL	Platinum
IR	Iron	NS	Nickel Silver
MS	Mild Steel	OR	Other/Unknown

Check with your local state agency for certification of compliance on Legal-for-Trade items. The weight accuracy class is referenced in the Description of Weights. Unless otherwise noted, the weights calibrated meet the requirements of the accuracy class. Results relate only to weights calibrated. The Surface Finishes of weights are evaluated visually. Weights are screened for magnetism using work instruction WI05-0035 when they are new, when requested by the customer or when weights are suspected of not meeting specifications. Density if measured is measured using OIML R111-1 (2004) method A2. Conventional Mass is reported based on a reference density of 8.0 g/cm³. The Uncertainty of Measurement is included in the determination of Maximum Permissible Error (MPE) Pass/Fail Criteria. The specifications for Maximum Permissible Error (MPE) can be found in NIST Handbook 105-1 (2019), NIST Handbook 105-1 (1990), ASTM E617-18 or OIML R111-1 (2004), manufacturer specifications or customer specifications.

The Uncertainty assigned to the Conventional Mass values are the result of the root-sum-square of the type A and type B components, calculated in accordance with NIST SOP 29 and the Guide to the expression of uncertainty in measurement, with coverage factor ($k=2$), to express the expanded uncertainty with an approximate 95.45% confidence level. This report is not to be used to claim product certification, approval, or endorsement by NVLAP, NIST, A2LA, or any government agency. This document and all data within, shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems.


Dan Demers, Metrologist

30 Sep 2022
Issued Date:

Prepared By:
9 Rice Lake Weighing Systems®●PN 64784●12/21
230 West Coleman Street●Rice Lake, WI 54868●USA
TEL: 715-234-9171 ●FAX: 715-234-6967
Definitions: <http://certs.ricelake.com/certs/DefinitionsV2.docx>
Page 1 of 2

Procedure used:
Internationally published procedures defined by NIST, ASTM and OIML

The Certificate of Weight Calibration (non-accredited) includes the following information:

1 Traceable certificate number	11 True Mass (Mass in Vacuum)	19 Record of the weighing instrument(s)
2 Contractor (sold to) name and address	12 True Mass Correction ⁰	20 Reference standard set used to calibrate item(s) listed on certificate
3 End user name and address	13 Conventional Mass: mass of a weight of a density of 8000 kg/m ³ which it balances in air of density of 1.2 kg/m ³	<i>0 The True Mass Correction is the deviation from the Nominal Value, reported in milligrams. A minus sign indicates that the True Mass of the weight is less than the nominal value.</i>
4 Date calibrated	14 Conventional Mass Correction ¹	<i>1 The Conventional Mass Correction is the deviation from the Nominal Value, reported in milligrams. A minus sign indicates that the Conventional Mass of the weight is less than the nominal value.</i>
5 Recalibration date (if requested)	15 A stated quantity of the estimated value of uncertainty ²	<i>2 All measurements have a degree of uncertainty regardless of precision and accuracy. This is caused by two factors, the limitation of the measuring instrument (systematic error) and the skill of the experimenter making the measurements (random error).</i>
6 NIST Certificate of Calibration Number	16 Maximum Permissible Error for the specified accuracy class	
7 Procedure used (Intercomparison Method)	17 Assumed material density of the weight being calibrated	
8 Identification of the calibrated item(s) and serial number, if applicable	18 Environmental conditions to time of calibration	
9 Name and address of the calibration laboratory		
10 Nominal mass		

Certificate of Weight Calibration															
RICE LAKE															
Traceable Certificate Number:		1234567		Temperature Range:		18		20.62 °C to 20.81 °C							
Client:		Rice Lake Weighing Systems		Pressure Range:		18		736.58 mmHg to 740.90 mmHg							
Date Calibrated:		26 Sep 2022 to 27 Sep 2022		Relative Humidity Range:		18		42 % to 51 %							
As Left Data (As Found Data is undifferentiated from As Left Data unless listed in As Found Data table)															
10	11	12	13	14	15	16	17	18	19	20					
Nominal Value	Unique ID	True Mass (Same UOM as Nom.)	True Mass Corr. (mg)	Conv. Mass (Same UOM as Nom.)	Conv. Mass Corr. (mg)	(k=2) Unc. (± mg)	MPE (± mg)	Pass (Y=Pass N=Fail)	Assumed Density (g/cm ³)	Assumed Material	Const. Type	Balance Used	Reference Standard Set Used	Air Density (mg/cm ³)	Clean Level
2 mg		2.00162	0.00162	2.00161	0.00161	0.00062	0.010	Y	7.95	SS	I	1605Q	K594Q	1.1659	A
3 mg		2.99858	-0.00142	2.99858	-0.00142	0.00062	0.010	Y	7.95	SS	I	1605Q	K594Q	1.1659	A
5 mg		5.00303	0.00303	5.00302	0.00302	0.00071	0.010	Y	7.95	SS	I	1605Q	K594Q	1.1660	A
10 mg		10.00376	0.00376	10.00375	0.00375	0.00097	0.010	Y	7.95	SS	I	1605Q	K594Q	1.1661	A
20 mg		20.00359	0.00359	20.00357	0.00357	0.00081	0.010	Y	7.95	SS	I	1605Q	K594Q	1.1661	A
30 mg		30.00011	0.00011	30.00008	0.00008	0.00081	0.010	Y	7.95	SS	I	1605Q	K594Q	1.1661	A
50 mg		49.9991	-0.0009	49.9990	-0.0010	0.0016	0.010	Y	7.95	SS	I	1605Q	K594Q	1.1662	A
100 mg		100.0033	0.0033	100.0032	0.0032	0.0019	0.010	Y	7.95	SS	I	1605Q	K594Q	1.1662	A
200 mg		200.0016	0.0016	200.0014	0.0014	0.0019	0.010	Y	7.95	SS	I	1605Q	K594Q	1.1663	A
300 mg		300.0032	0.0032	300.0029	0.0029	0.0016	0.010	Y	7.95	SS	I	1605Q	K594Q	1.1664	A
500 mg		500.0026	0.0026	500.0021	0.0021	0.0021	0.010	Y	7.95	SS	I	1605Q	K594Q	1.1664	A
1 g		1.0000130	0.0130	1.0000121	0.0121	0.0027	0.034	Y	7.95	SS	I	1605Q	K594Q	1.1663	A
2 g		1.9999822	-0.0178	1.9999803	-0.0197	0.0031	0.034	Y	7.95	SS	I	1605Q	K594Q	1.1662	A
3 g		3.0000197	0.0197	3.0000168	0.0168	0.0033	0.034	Y	7.95	SS	I	1605Q	K594Q	1.1663	A
5 g		5.0000101	0.0101	5.0000054	0.0054	0.0055	0.034	Y	7.95	SS	I	1605Q	K594Q	1.1664	A
10 g		10.000031	0.031	10.000021	0.021	0.011	0.050	Y	7.95	SS	II	676Q	K594Q	1.1664	A
20 g		20.000042	0.042	20.000023	0.023	0.012	0.074	Y	7.95	SS	II	676Q	K594Q	1.1663	A
30 g		30.000008	0.008	29.999980	-0.020	0.019	0.074	Y	7.95	SS	II	1631Q	K594Q	1.1659	A
50 g		49.999981	-0.019	49.999934	-0.066	0.027	0.12	Y	7.95	SS	II	1631Q	K594Q	1.1657	A
★ 100 g		100.000281	0.281	100.000187	0.187	0.052	0.25	Y	7.95	SS	II	1631Q	K594Q	1.1595	A
As Found Data															
Nominal Value	Unique ID	True Mass (Same UOM as Nom.)	True Mass Corr. (mg)	Conv. Mass (Same UOM as Nom.)	Conv. Mass Corr. (mg)	(k=2) Unc. (± mg)	MPE (± mg)	Pass (Y=Pass N=Fail)	Assumed Density (g/cm ³)	Assumed Material	Const. Type	Balance Used	Reference Standard Set Used	Air Density (mg/cm ³)	Clean Level
★ 100 g		99.999884	-0.116	99.999789	-0.211	0.052	0.25	N X	7.95	SS	II	1631Q	K594Q	1.1658	A